

Aditya Dispomed Products Pvt. Ltd. Plot No 19, Sector 6 IMT Manesar, Haryana 122050 India T: +91 124 4764900

www.adityadispomed.com

Document Revision Date IFU 01 16

15-02-2025

INSTRUCTIONS FOR USE STERILE SURGICAL BLADES

DESCRIPTION:

Surgical Blades are made of Carbon steel / Stainless Steel. They are packed in Peel Apart foil pouches. Each blade is individually packed in a foil pack to prevent the edge from getting damaged in contact with another blade. Carbon steel have a protective VCI liner. Device is intended for single use and is sterilized by Gamma Irradiation.

Labelling and Marking of blades is fully compliant with the requirements of MDD93/42/EEC, EU MDR 2017/745, EN ISO 20417:2021, BS 2982:1992 & EN ISO 15223-1:2021.

INTENDED USE:

These are used for surgery after mounting on Metal B.P. Handles.

INDICATION:

- Dissection incision
- Different size for different thickness of tissue and site.
- Removal of extra tissues from operating area.

INTENDED USER

This device must be used by surgeon, doctor or paramedic staff who are trained and qualified in the appropriate surgical techniques.

PATIENT TARGET GROUP

Device can be used for all age patients group based on decision by surgeon, doctor or paramedic staff.

CONTRAINDICATIONS:

- Re-use of blade can work as carrier for communicable disease to patient and/or user
- Adverse event may be happen if blade use after expiry date as expiry date of product sterility

STORAGE CONDITION:

- Keep away from direct sun light.
- Keep away from rain
- Storage temperature should be 5°C to 45°C
- · Keep away from children
- Store in cool and dry place.

SHELF-LIFE:

• 5 Years (60 months) from the date of manufacturing.

PRECAUTIONS:

- This device must only be used by surgeons who are trained & qualified in the appropriate skills.
- Care must be taken so that the pouch is not opened in an unsterile area otherwise the blade which has already been sterilized by Gamma radiation will become unsterile.
- Don't use if product or pouch appear to be damaged
- For single patient use only content supplied sterile
- · During use avoid twisting, bending or putting excessive force or strain on the blade in order to prevent breakage.
- Always open the pouch from peel apart direction to avoid injury.
- Devices are extremely sharp, use care while handling.
- Proper procedures must be used as applicable for handling any sterile product.
- Care must be taken during disposal of device to avoid any contact or injury due to the sharp nature of the device.
- · Proper care must be taken while fitment of blade with respective handle to avoid any injury or accident
- In case of changes in the performance of device for intended use replace the defective device by new to full fill the required application

Page **1** of **4**







Aditya Dispomed Products Pvt. Ltd. Plot No 19, Sector 6 IMT Manesar, Haryana 122050 India T: +91 124 4764900

www.adityadispomed.com

Document Revision Date IFU 01 16

15-02-2025

USE

- It is recommended that the foil pouch is peeled apart half in a sterile area allowing the blade to be exposed. There is no need to drop the blade from the pack.
- Blade shall be used with Fitment# 3 and Fitment# 4 handles conforming to EN ISO 27740.
- The blade can be either directly fitted on to metal handle by sliding the handle in the slot made on the blade before removing the blade from the pouch or the blade can removed from the pouch with a forceps and fitted on the handle as shown in the attached sketch.

DISPOSAL: Discard the Surgical Blades in proper waste container & dispose off the product in accordance with accepted medical practice and applicable local, state and country laws and regulations for handling of bio-medical waste



WARNING

- Use according to the instruction
- The product should be used only by a qualified surgeon, Doctor or paramedic.
- Before use always check integrity of product and packing along with expiry date.
- For single use only,
- Do not reuse the device in any circumstances
- Do not Resterile the device
- Use product immediately after opening the pack
- ADPPL is not responsible for any possible consequences resulting from improper use.
- ADPPL do not hold any responsibility if device re-used or re-sterile.
- Sterility of product is not guaranteed if packet is broken/torn.
- Keep out of reach of children.
- After use of products must be disposed off as per country law of bio-waste handling rule

KNOWN CHARACTERISTICS OF DEVICE & TECHNICAL FACTOR FOR THE RISK IN CASE OF RE-USE

- Difficult to cut at incision site during reuse.
- Any infectious disease can transfer.
- In case of reuse of device may cause communicable and transferable disease from the device.

RETURN OF DEVICE

The documents & papers required for return of defective device and in compliance of law of the land for origin and destination shall be provided to ADPPL as soon as possible after receipt of the product along with the photograph & inspection report as an evidence of damaged product and the product shall not be used under any circumstances & it should be separated with proper identification of damaged goods.

NOTICE TO THE USER

Any serious incident that occurs in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established. The details for the notice to the manufacturer of any serious incident can be reported through our web site www.adityadispomed.com and customer complaint contact T: +91 124 4764900.

In case of reporting to regulatory authority/competent authority of the member state as per law of the land for reporting of such incidents as per the regulation where the user is established.





Schiffgraben 41 30175 Hannover,

MDSS GmbH



Aditya Dispomed Products Pvt. Ltd. Plot No 19, Sector 6 IMT Manesar, Haryana 122050 India

T: +91 124 4764900

www.adityadispomed.com

Document Revision Date IFU 01 16

15-02-2025

SYMBOLS

LOT	Batch code	REF	Catalogue number	\subseteq	Use by date
②	Do not re-use	\triangle	Caution, consult instruction for use	®	Do not use if package is damaged and consult instructions for use
*	Keep dry	*	Keep away from Sunlight	STERRIUZE	Do not resterilize
STERILE R	Sterilized using Irradiation	5°C	Temperature Limitation	w	Manufacturer
Ţi	Consult instruction for use	EC REP	Authorized representative in the European Community	(€ 2 4 6 0	CE Mark with NB Number
×	Non-pyrogenic	_\	Country of Manufacturer	MD	Medical Device

The same IFU is available on our web site





Aditya Dispomed Products Pvt. Ltd. Plot No 19, Sector 6 IMT Manesar, Haryana 122050 India T: +91 124 4764900

www.adityadispomed.com

Document IFU 01 Revision 16

Date 15-02-2025

RECOMMENDED PROCEDURE FOR PEELING OFF BLADE PACKING, ATTACHMENT & REMOVAL OF BLADES

PEEL OFF BLADE PACKING

1- PACKED SAMPLE



2- HOLD IT IN HAND



3- PEEL OPEN POSITION PUT FORCE HERE TO OPEN HOLD IT WITH THUMB

4-POSITION OF BLADE



5. HOLD IT IN THIS POSITION TO AVOID FALLING OF BLADE



ATTACHMENT & REMOVAL OF BLADES

To attach blade

Grip blade with forceps, or similar, avoiding contact with cutting edge.

- Hold handle in left hand with fitting uppermost (fig 1).
- Place blade partway over handle fitting and engage slots (fig 2).
- Slide blade until it clicks into position. (fig 3).






To remove a blade

Grip blade with forceps, or similar.

Lift heel of blade at point 'A' with tip of index finger, avoiding contact with cutting edge, and slide away from handle carefully (fig 4).



DISPOSAL: Discard the Surgical Blades in proper waste container & dispose off the product in accordance with accepted medical practice and applicable local, State and Country laws and regulations for handling of bio-medical waste.

Page 4 of 4